NOTES:

1. SUBSTRATE MATERIAL: N-BK7

2. COATING (APPLY ACROSS CLEAR APERTURE)

S1 & S2: VIS-NIR

R(abs) ≤0.25% @ 880nm

R(avg) ≤1.25% @ 400 - 870nm


R(avg) ≤1.25% @ 890 - 1000nm




4. POWER, IRREGULARITY, AND SURFACE QUALITY SPECIFICAIONS APPLY ACROSS CLEAR APERTURE

5. Rohs Compliant







SPECIFICATIONS SUBJECT TO CHANGE WITHOUT NOTICE DIMENSIONS ARE FOR REFERENCE ONLY

REV. A	S1	S2					
SHAPE	CONVEX	PLANO	EFL	-50		Edmund Ontion	
RADIUS	25.84	∞	BFL	-51.65	Ul	Edmund Optics®	
SURFACE QUALITY	60-40	60-40		1		12.5mm H X 25mm L X -50mm FL VIS-NIR	
CLEAR APERTURE	11.50 X 24.00	11.50 X 24.00	THIRD ANGLE PROJECTION	\bigoplus	TITLE	COATED, CYLINDER LENS	
COATING APERTURE	11.50 X 24.00	11.50 X 24.00				·	
BEVEL	PROTECTED AS NEEDED	PROTECTED AS NEEDED	ALL DIMS IN	mm	DWG NO	69839 SHEE	